



DECLARATION OF CONFORMITY

ACCORDING TO (EU) 2017/745 MEDICAL DEVICE REGULATION

EU Representative

Name: SUNGO EUROPE B.V.
Address: Fascinatio Boulevard 522, Unit 1.7,
2909VA Capelle aan den IJssel, The Netherlands
SRN: NL-AR-000000247

Manufacturer

Name: Ziyang Freqty Medical Equipment Co., Ltd.
Address: Floor 2-3, unit 7, building 3, No. 222,
West Section 3, outer ring road, Yanjiang District,
Ziyang City, Sichuan Province, China
SRN: CN-MF-000019421

Conformity Assessment

Conformity Assessment Procedure

Annex II+III of Regulation (EU) 2017/745
RoHS 2.0 Directive 2011/65/EU & (EU) 2015/863
WEEE Directive 2012/19/EU

Applicable Standards

ISO 13485:2016 EN ISO 14971:2019
ISO 15223-1:2021 ISO 20417:2021
IEC 60601-1:2005/AMD2:2020
IEC 60601-1-2:2014+A1:2020
IEC 60825-1:2014 IEC 62471:2006
IEC 60601-1-6:2010+AMD1:2013+AMD2:2020 CSV
IEC 62366-1:2015/AMD1:2020
IEC 62304:2006/AMD1:2015
ISO 10993-1:2018 ISO 10993-5:2009
ISO 10993-10:2021 ISO 10993-23:2021

Remark

The declaration of conformity is valid in connection with the release technical document ZYPT-CE-17.
All the supporting documentation is retained at the premises of the manufacturer.
The Declaration of Conformity is exclusively under the sole responsibility of the manufacturer.

Product Information

Name: Intraoral Digital Impression Instrument
Model: PANDA free
GMDN: 63669
EMDN: Z11030402
Basic UDI-DI: 697513623PTPANDA001M9
Classification: Class I
Rule: Rule 5, Rule 13, Annex VIII, Regulation (EU) 2017/745

Declaration

We herewith declare that the above-mentioned products meet the requirements of Medical Device Regulation (EU) 2017/745 and the applicable standards above.

Signature:  Date: 2025.04.14
Ao Mingwu

Position: General Manager Place: Ziyang / China